

CERTIFICATE OF CALIBRATION

Certificate No.: 11641

Date: 30/05/2006

Next Calibration Due: 29/11/2006

Customer: Aftercrete Constructional Engineering Co Ltd
5 Nursery Road

Luton
LU3 2RG

Manufacturer: BW

Model No.: GAS ALERT MICRO

Serial No.: 5353763737

Asset No.: NONE

GAS TYPES	CONCENTRATION	CYLINDER REFERENCE
OXYGEN	20.9 % VOLUME	AMBIENT
METHANE	50 % LEL	2407
HYDROGEN SULPHIDE	70 PPM	2478
CARBON MONOXIDE	100 PPM	7660

Parts Fitted

Filter(s)

Oxygen Sensor

This is to certify that the instrument detailed above has been calibrated and performance checked to meet the required standard.

Signature.....

Sure-Safe Ltd
College Farm Business Centre
North End, Meldreth
Royston, Herts
SG8 6NT

Tel: 01763 262649
Fax: 01763 263134

Service@sure-safe.com

Gas Detection at its Best